

## LoRa® SPY T1



Temperature and door-opening recorder with a long-range LoRaWAN $^{\text{TM}}$  connectivity



on contractual photo

## **Presentation**

The LoRa® SPYT1 recorder measures and records temperature and door openings. Data are sent to the JRI secure Cloud by the LoRaWAN $^{\text{TM}}$  network of the telecom operators members of the LoRa Alliance $^{\text{TM}}$ , or by a local network. In case of thresholds excursion, real-time alerts are sent.

The LoRa® SPYT1 is particularly adapted to the supply chain applications, transport and warehouse storage areas.

## Technical features

Interface	Status LED lights, On/Off touch sensitive button, LCD display
Measurement range	-30°C to +70°C
Accuracy	$\pm 0.4^{\circ}\text{C}$ from -20°C to +40°C and $\pm 0.5^{\circ}\text{C}$ out of this range
Memory	10 000 data points
Communication frequency	868 MHz
Radio range	Up to 16 km open field range
Sensors	Class A PT 100 and light detector
Resolution	0.01
Measuring interval	Adjustable from 1 mn to 12 hrs
Transmission & recording interval (internal memory)	Adjustable from 1 mn to 24 hrs
Response time $\sigma_{\scriptscriptstyle{90}}$	~15 mn
Operating conditions	-30°C to +70°C
Protection	IP68
Case	Polycarbonate – Food contact
Dimensions and weight	87 x 64 x 25 mm / ~ 100 g
Power supply	Replaceable Lithium 3,6v battery
Battery lifetime	2 to 4 years depending on use
Fixation	Fixing eyelets and integrated magnets
Supplied with	Battery, Grey identification ring User manual on www.jri-maxant.com
Compliance	CE, ROHS, FCC, ETSI 300-328, NF EN 12830
Standard calibration points	-18°C/+5°C/+25°C

JRI, Société par actions simplifiée au capital de 4 000 000 €

Pôle logistique / 2 Rue de la Voivre / PA Technoland / BP 21 / 25490 FESCHES LE CHÂTEL / France SIRET 380 332 858 00030 - Tél : +33 (0)3 81 30 68 04 / Fax : +33 (0)3 81 30 60 99 / sales@jri.fr


## **Benefits**

Easy-to-use
 The recorder is ready-to-use and

has operating and alarm lights



- Excellent measurement accuracy
- Several modes of data exploitation





PART NRS 11802 : LoRa® SPY T1 11954 : LoRa® SPY T1 SMART 17 348b EN 01 02

www.jri-maxant.com